



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

International Light Technologies
10 Technology Drive, Peabody, MA 01960

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025:2017
& Meets the Requirements of ANSI/NCSI Z540.1-1994
& ANSI/NCSI Z540.3-2006 subclause 5.3

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

Calibration of Electrical and Optical Devices
(As detailed in the supplement)

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen
President

Initial Accreditation Date:

January 02, 2012

Issue Date:

March 04, 2024

Expiration Date:

May 31, 2026

Accreditation No.:

66765

Certificate No.:

L24-181-1

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: www.pjlabs.com



Certificate of Accreditation: Supplement

International Light Technologies

10 Technology Drive, Peabody, MA 01960

Contact Name: John Ellis Phone: 407-961-6383

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Research Radiometers & Photometers Fixed points ^F	1 mA	3.6×10^{-7} A	Keithley Current Calibrator	TP-0112 TP-0113 TP-0123 TP-0135 TP-0116
	100 uA	3.4×10^{-8} A		
	10 uA	6.2×10^{-9} A		
	1 uA	6.5×10^{-10} A		
	100 nA	7.3×10^{-11} A		
	10 nA	7.9×10^{-12} A		
	1 nA	8.4×10^{-13} A		
	100 pA	2.6×10^{-13} A		
	10 pA	1.3×10^{-13} A		
1 pA	4.2×10^{-14} A			

Optical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Spectroradiometers Spectral Irradiance At the listed Wavelengths ^F			Calibrated Standard Lamp	OP-0152
200 nm to 350 nm	3×10^{-7} W/cm ² nm to 1×10^{-6} W/cm ² nm	4 % of Reading		
350 nm to 400 nm	4×10^{-8} W/cm ² nm to 2×10^{-6} W/cm ² nm	3 % of Reading		
400 nm to 900 nm	6×10^{-6} W/cm ² nm to 2×10^{-5} W/cm ² nm	3 % of Reading		
900 nm to 1 050 nm	6×10^{-6} W/cm ² nm to 2×10^{-5} W/cm ² nm	7 % of Reading		
1 050 nm to 1 250 nm	6×10^{-6} W/cm ² nm to 2×10^{-5} W/cm ² nm	15 % of Reading		
1 250 nm to 2 050 nm	5×10^{-6} W/cm ² nm to 1.5×10^{-5} W/cm ² nm	7 % of Reading		
1 250 nm to 2 500 nm	2×10^{-6} W/cm ² nm to 6×10^{-6} W/cm ² nm	14 % of Reading		
Spectroradiometers Spectral Radiance At the listed Wavelengths ^F				
350 nm to 400 nm	2×10^{-7} W/sr cm ² nm to 6×10^{-7} W/sr cm ² nm	3 % of Reading		
400 nm to 900 nm	2×10^{-6} W/sr cm ² nm to 7×10^{-6} W/sr cm ² nm	3 % of Reading		
900 nm to 1 050 nm	2×10^{-6} W/sr cm ² nm to 7×10^{-6} W/sr cm ² nm	7 % of Reading		



Certificate of Accreditation: Supplement

International Light Technologies

10 Technology Drive, Peabody, MA 01960

Contact Name: John Ellis Phone: 407-961-6383

Accreditation is granted to the facility to perform the following calibrations:

Optical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Spectroradiometers Spectral flux At the listed Wavelengths ^F			Calibrated Standard Lamp	OP-0152
200 nm to 350 nm	1×10^{-7} W/nm to 1×10^{-5} W/nm	4 % of Reading		
350 nm to 400 nm	2×10^{-7} W/nm to 1×10^{-5} W/nm	3 % of Reading		
400 nm to 900 nm	2.5×10^{-6} W/nm to 2.5×10^{-4} W/m	3 % of Reading		
900 nm to 1 050 nm	2.5×10^{-6} W/nm to 2.5×10^{-4} W/m	7 % of Reading		
1 050 nm to 1 250 nm	3×10^{-6} W/nm to 2.5×10^{-4} W/m	15 % of Reading		
1 250 nm to 2 050 nm	2×10^{-6} W/nm to 2×10^{-4} W/m	7 % of Reading		
1 250 nm to 2 500 nm	7×10^{-7} W/nm to 8×10^{-5} W/m	14 % of Reading		
Belt Radiometers 200 nm to 500 nm	10 mW/cm ² to 200 mW/cm ²	6.2 % of Reading	Calibrated Standard Lamp	OP-0054, OP-0044, OP-0055, OP-0035, OP-0053, OP-0119, OP-118, OP-0121, OP-0124, OP-0125, OP-120, OP-0042, OP-0013, OP-0043
Handheld Radiometers At the listed Wavelengths ^F			Silicon Photodiodes Phototubes	OP-0030
200 nm to 250 nm	1×10^{-10} W/cm ² to 8×10^{-6} W/cm ²	11 % of Reading		
250 nm to 400 nm	1×10^{-10} W/cm ² to 7×10^{-3} W/cm ²	5 % of Reading		
400 nm to 960 nm	1×10^{-10} W/cm ² to 2×10^{-2} W/cm ²	4 % of Reading		
Extended UV Scanned Irradiance At the listed Wavelengths ^F				OP-0036
200 nm to 250 nm	1×10^{-10} W/cm ² to 8×10^{-6} W/cm ²	7.1 % of Reading		
250 nm to 400 nm	1×10^{-10} W/cm ² to 7×10^{-3} W/cm ²	4.1 % of Reading		



Certificate of Accreditation: Supplement

International Light Technologies

10 Technology Drive, Peabody, MA 01960
 Contact Name: John Ellis Phone: 407-961-6383

Accreditation is granted to the facility to perform the following calibrations:

Optical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Irradiance Response At the listed Wavelengths ^F			Silicon Photodiodes Phototubes	OP-0131
200 nm to 250 nm	1×10^{-10} W/cm ² to 8×10^{-6} W/cm ²	11 % of Reading		
250 nm to 400 nm	1×10^{-10} W/cm ² to 7×10^{-3} W/cm ²	4.5 % of Reading		OP-0131, OP-0018, OP-0098, OP-0087, OP-0017, OP-0096, OP-0019, OP-0007, OP-0029, OP-0131
400 nm to 960 nm	1×10^{-10} W/cm ² to 2×10^{-2} W/cm ²	3 % of Reading		OP-0029, OP-0018, OP-0131
960 nm to 1 000 nm	1×10^{-10} W/cm ² to 1×10^{-4} W/cm ²	4.5 % of Reading		OP-0029, OP-0131, OP-0002
1 000 nm to 1 100 nm	1×10^{-10} W/cm ² to 1×10^{-4} W/cm ²	5 % of Reading		OP-0002, OP-0018, OP-0131
Radiance Response At the listed Wavelengths ^F				OP-0041
200 nm to 250 nm	3×10^{-11} W/cm ² /sr to 3×10^{-6} W/cm ² /sr	11 % of Reading		
250 nm to 400 nm	3×10^{-11} W/cm ² /sr to 3×10^{-3} W/cm ² /sr	4.5 % of Reading		
400 nm to 960 nm	3×10^{-11} W/cm ² /sr to 7×10^{-3} W/cm ² /sr	3 % of Reading		
960 nm to 1 000 nm	3×10^{-11} W/cm ² /sr to 3×10^{-5} W/cm ² /sr	4.5 % of Reading		
1 000 nm to 1 100 nm	3×10^{-11} W/cm ² /sr to 3×10^{-5} W/cm ² /sr	5 % of Reading		



Certificate of Accreditation: Supplement

International Light Technologies

10 Technology Drive, Peabody, MA 01960

Contact Name: John Ellis Phone: 407-961-6383

Accreditation is granted to the facility to perform the following calibrations:

Optical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED	
Radiant Intensity Response Based on 125 mm distance. Values vary per application distance. At the listed Wavelengths ^F			Silicon Photodiode	OP-0037	
200 nm to 250 nm	1 x 10 ⁻⁸ W/sr to 2 x 10 ⁻³ W/sr	11 % of Reading			
250 nm to 400 nm	1 x 10 ⁻⁸ W/sr to 1 W/sr	4.5 % of Reading			
400 nm to 960 nm	1 x 10 ⁻⁸ W/sr to 3 W/sr	3 % of Reading			
960 nm to 1 000 nm	1 x 10 ⁻⁸ W/sr to 2 x 10 ⁻² W/sr	4.5 % of Reading			
1 000 nm to 1 100 nm	1 x 10 ⁻⁸ W/sr to 2 x 10 ⁻² W/sr	5 % of Reading			
Radiant Power Response At the listed Wavelengths ^F					OP-0049
200 nm to 250 nm	4 x 10 ⁻¹² W to 4 x 10 ⁻⁷ W	11 % of Reading			OP-0106
250 nm to 400 nm	4 x 10 ⁻¹² W to 3 x 10 ⁻⁴ W	4.5 % of Reading			
960 nm to 1 000 nm	4 x 10 ⁻¹² W to 4 x 10 ⁻⁶ W	4.5 % of Reading			
1 000 nm to 1 100 nm	4 x 10 ⁻¹² W to 4 x 10 ⁻⁶ W	5 % of Reading			
400 nm to 960 nm	4 x 10 ⁻¹² W to 3 x 10 ⁻³ W	3 % of Reading			
Illuminance Sensitivity At the listed Wavelength ^F					OP-0070
400 nm to 700 nm	9 x 10 ⁻⁴ lx to 20 klx	2.4 % of Reading			OP-0071
Luminance Sensitivity At the listed Wavelengths ^F					
400 nm to 700 nm	2 x 10 ⁻⁴ cd/m ² to 60 kcd/m ²	2.4 % of Reading	OP-0081		
Luminous Intensity of a Standard Lamp At the listed Wavelengths ^F			OP-0025		
400 nm to 700 nm	3 x 10 ⁻⁷ cd to 4 x 10 ⁺² cd	2.4 % of Reading			
Luminous Intensity Sensitivity Based on 125 mm distance. Values vary per application distance. At the listed Wavelengths ^F			OP-0072		
400 nm to 700 nm	1 x 10 ⁻⁷ cd to 30 cd	2.4 % of Reading			
Luminous Power Sensitivity At the listed Wavelengths ^F			OP-0133		
400 nm to 700 nm	3 x 10 ⁻⁸ lm to 8 lm	2.4 % of Reading			
Filter Transmission			Agilent UV-VIS Spectroscopy		
200 nm to 1 000 nm	1 % to 100 %	+/- 0.1 % of Reading			



Certificate of Accreditation: Supplement

International Light Technologies

10 Technology Drive, Peabody, MA 01960

Contact Name: John Ellis Phone: 407-961-6383

Accreditation is granted to the facility to perform the following calibrations:

Optical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE (AND SPECIFICATION WHERE APPROPRIATE)	CALIBRATION OR MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	CALIBRATION MEASUREMENT METHOD OR PROCEDURES USED
Optical Density ^F			Agilent UV-VIS Spectroscopy	OP-0133
200 nm to 1 000 nm	Up to 2.5 OD	+/- 0.005 OD of Reading		
Wavelength				
200 nm to 1 000 nm	200 nm to 1 000 nm	+/- 0.5 nm of Reading		
Degrees Kelvin ^F			ILT950/960 Spectrometer	OP-0155
Correlated Color Temperature (CCT) of Meters and Sources	2 000 K - 6 000 K	1 % of Reading		
	6 001 K - 7 500K	1.2 % of Reading		
	7 501 K - 10 000K	1.7 % of Reading		

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor k (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
3. The presence of a superscript F means that the laboratory performs calibration of the indicated parameter at its fixed location.
4. This is the primary site for all quality management system activities.